


SPECTRUM DATA

QB12

DATE : Feb. 08. 2007

Wavelength(nm)	400	410	420	430	440	450	460	470	480	490	500
Transmissivity(%)	69.03	72.44	75.07	76.88	78.63	80.33	81.00	81.19	80.85	79.25	76.17
Wavelength(nm)	510	520	530	540	550	560	570	580	590	600	610
Transmissivity(%)	71.52	65.10	57.46	48.92	39.82	31.19	24.10	17.82	12.84	9.15	6.40
Wavelength(nm)	620	630	640	650	660	670	680	690	700	710	720
Transmissivity(%)	4.34	2.96	2.05	1.40	0.97	0.95	0.98	0.38	0.30	0.24	0.20



Color Glass Filter (OPTIMA INC. JAPAN)

- 1) Model : Abs. Calibration Filter QB12
- 2) Test Instrument : SP-3000 Plus
- 3) Scan Interval : 2 nm
- 4) Part No.: 30010

Inspector : Technomart Inc. Eng. Div.

SEO, IL-SOO

]